



# 150 ANOS PARA 118 ELEMENTOS **A TABELA PERIÓDICA**

## **XXVI ENCONTRO NACIONAL** DA SOCIEDADE PORTUGUESA DE QUÍMICA

24, 25 E 26 DE JULHO DE 2019  
FACULDADE DE CIÊNCIAS DA UNIVERSIDADE DO PORTO



ANO INTERNACIONAL  
DA TABELA PERIÓDICA



## Quantification of salivary magnesium using a microfluidic paper based analytical device ( $\mu$ PAD)

Mafalda T. S. Silva, Juliana I. S. Aguiar, Raquel B. R. Mesquita, António O. S. S. Rangel

Universidade Católica Portuguesa, CBQF – Centro de Biotecnologia e Química Fina – Laboratório Associado,  
Escola Superior de Biotecnologia, 4169-005 Porto, Portugal

Email: [rmesquita@porto.ucp.pt](mailto:rmesquita@porto.ucp.pt)

Magnesium is a mineral present in the human body involved in several physiological processes, namely acting as a cofactor in more than 300 enzymatic reactions. It allows the proper functioning of the enzymes enabling several reactions in the body. In this context, due to the importance of this mineral for human health, its concentration for each tissue/fluid must be monitored; otherwise, there could be a case of magnesium deficit (hypomagnesemia), or excess of magnesium (hypermagnesemia).<sup>1</sup>

The magnesium cation is present in the saliva at about 0.2 mM. However, some studies show that certain pathologies/disturbances, namely cystic fibrosis and burning mouth syndrome,<sup>2</sup> may be associated with alterations in the concentration of the salivary magnesium. Thus, its quantification in saliva can be an indicator of those pathologies.

Microfluidic paper-based analytical devices ( $\mu$ PADs) have several benefits that make them ideally suited to conduct in-situ, on-hand determinations. While many of the current monitoring techniques require specialised skills, laborious laboratory processes, or/and expensive equipment,  $\mu$ PADs provide a quick and cheap analytical result. Additionally, the low-cost, easy-disposable and intrinsic portability of the  $\mu$ PADs, make them ideal devices for regular monitoring i.e. routine consultation.

In this context, the objective of described work was to design a  $\mu$ PAD for a fast, in-the moment, reliable magnesium quantification in human saliva (**Figure 1**). The method was based on the reaction between magnesium and eriochrome cyanine dye, generating a coloured complex, quantified through pixel readings. The overall consumption of both reagents and sample is rather low. The developed  $\mu$ PAD enables to determine the salivary magnesium concentration as an economical and in-situ alternative to be used by unskilled operators.




Figure 1: Schematic representation of developed  $\mu$ PAD: a) exploded view of the assembly; b) colour scheme of the calibration curve.

Acknowledgements: This work was supported by National Funds from Fundação para a Ciência e a Tecnologia (FCT) through project UID/Multi/50016/2019.

### References:

1. Jahnhen-Dechent W.; Kettelet M. *Clin. Kidney J.* **2012**, 5, i3.
2. a) Livnat G.; Bentur L.; Kuzmisky E.; Nagler R. *J. Oral Pathol. Med.* **2010**, 39, 16. b) Henkin R.; Gouliouk V.; Fordyce A. *Arch. Oral Biol.* **2012**, 57, 205.